

Facility Name _____
 Month / Year: _____

Laboratory Temperatures Data

Date	Time	Analyst	Ambient	Refrigerator #1	Refrigerator #2	BOD Incubator	Bacteriology Incubator	Bacteriology Water Bath	Drying Oven #1	Drying Oven #2	Muffle Furnace	Autosampler #1	Autosampler #2
Acceptable Range			~ 20°C	≤ 6°C	≤ 6°C	20 ± 1°C	35 ± 0.5°C	44.5 ± 0.2°C	104 ± 1°C	104 ± 1°C	550°C	≤ 6°C	≤ 6°C
1													
2													
3													
4													
5													
6													
7													
8													
9													
10													
11													
12													
13													
14													
15													
16													
17													
18													
19													
20													
21													
22													
23													
24													
25													
26													
27													
28													
29													
30													
31													